

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031022\
 Data File : VW024915.D
 Acq On : 10 Mar 2022 14:57
 Operator : SY/MD
 Sample : N1855-16
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D44

Quant Time: Mar 11 02:48:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 02:45:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	289377	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	281813	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	126269	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	127023	43.170	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.340%
7) Chloroethane-d5	1.568	69	98129	42.507	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.020%
11) 1,1-Dichloroethene-d2	2.108	63	162552	32.914	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.820%
21) 2-Butanone-d5	3.892	46	135677	81.024	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.020%
24) Chloroform-d	4.349	84	208311	40.556	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.120%
26) 1,2-Dichloroethane-d4	5.030	65	125492	41.408	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.820%
32) Benzene-d6	5.050	84	445110	42.612	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.220%
36) 1,2-Dichloropropane-d6	6.069	67	139095	42.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.480%
41) Toluene-d8	7.313	98	386492	41.765	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.540%
43) trans-1,3-Dichloroprop...	7.622	79	56082	39.659	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.320%
47) 2-Hexanone-d5	8.088	63	90575	73.323	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.320%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	159598	41.346	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.700%
66) 1,2-Dichlorobenzene-d4	11.622	152	126681	43.553	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.100%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.918	95	7577	2.906	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024915.D
Acq On : 10 Mar 2022 14:57
Operator : SY/MD
Sample : N1855-16
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D44

Quant Time: Mar 11 02:48:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 02:45:24 2022
Response via : Initial Calibration

