

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031022\
 Data File : VW024917.D
 Acq On : 10 Mar 2022 15:44
 Operator : SY/MD
 Sample : N1855-17
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D46

Quant Time: Mar 11 02:48:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 02:45:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	289013	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	282780	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	125927	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	127028	43.226	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.460%
7) Chloroethane-d5	1.568	69	98030	42.518	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.040%
11) 1,1-Dichloroethene-d2	2.108	63	172382	34.949	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.900%
21) 2-Butanone-d5	3.895	46	134643	80.508	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.510%
24) Chloroform-d	4.349	84	208086	40.563	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.120%
26) 1,2-Dichloroethane-d4	5.030	65	127790	42.220	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.440%
32) Benzene-d6	5.050	84	442406	42.209	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.420%
36) 1,2-Dichloropropane-d6	6.069	67	138149	42.305	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.600%
41) Toluene-d8	7.313	98	385773	41.545	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.100%
43) trans-1,3-Dichloroprop...	7.622	79	56446	39.780	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.560%
47) 2-Hexanone-d5	8.088	63	89853	72.489	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.490%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	158947	41.037	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.080%
66) 1,2-Dichlorobenzene-d4	11.622	152	129753	44.731	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.460%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.121	96	21926	10.249	ug/L	# 15
19) 1,1-Dichloroethane	3.198	63	3678	0.882	ug/L	98
20) cis-1,2-Dichloroethene	3.915	96	38657	15.436	ug/L	95
30) 1,1,1-Trichloroethane	4.616	97	9154	2.458	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024917.D
Acq On : 10 Mar 2022 15:44
Operator : SY/MD
Sample : N1855-17
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D46

Quant Time: Mar 11 02:48:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 02:45:24 2022
Response via : Initial Calibration

