

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024926.D
 Acq On : 10 Mar 2022 19:18
 Operator : SY/MD
 Sample : N1879-13
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DA6

Quant Time: Mar 11 02:50:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 02:45:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	292540	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	283814	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	126483	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	122188	41.077	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	82.160%		
7) Chloroethane-d5	1.568	69	94011	40.283	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	80.560%		
11) 1,1-Dichloroethene-d2	2.108	63	154615	30.969	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	61.940%		
21) 2-Butanone-d5	3.896	46	140332	82.898	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	82.900%		
24) Chloroform-d	4.349	84	207427	39.947	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	79.900%		
26) 1,2-Dichloroethane-d4	5.031	65	126661	41.342	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.680%		
32) Benzene-d6	5.050	84	437237	41.564	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.120%		
36) 1,2-Dichloropropane-d6	6.069	67	136983	41.795	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	83.580%		
41) Toluene-d8	7.314	98	378141	40.575	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	81.140%		
43) trans-1,3-Dichloroprop...	7.619	79	55039	38.647	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.300%		
47) 2-Hexanone-d5	8.088	63	96491	77.561	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	77.560%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	164527	42.322	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	84.640%		
66) 1,2-Dichlorobenzene-d4	11.622	152	126946	43.571	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.140%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	3.918	96	3519	1.388	ug/L	86
34) Trichloroethene	5.915	95	81421	31.009	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024926.D
Acq On : 10 Mar 2022 19:18
Operator : SY/MD
Sample : N1879-13
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DA6

Quant Time: Mar 11 02:50:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 02:45:24 2022
Response via : Initial Calibration

