

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024938.D
 Acq On : 11 Mar 2022 00:53
 Operator : SY/MD
 Sample : N1879-07
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D97

Quant Time: Mar 11 05:27:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	281651	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	275965	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	124653	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	114990	40.152	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.300%
7) Chloroethane-d5	1.571	69	89697	39.921	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.840%
11) 1,1-Dichloroethene-d2	2.108	63	147965	30.783	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.560%
21) 2-Butanone-d5	3.892	46	140056	85.933	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.930%
24) Chloroform-d	4.349	84	206284	41.263	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.520%
26) 1,2-Dichloroethane-d4	5.030	65	125157	42.430	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.860%
32) Benzene-d6	5.050	84	430192	42.057	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.120%
36) 1,2-Dichloropropane-d6	6.069	67	136264	42.758	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.520%
41) Toluene-d8	7.313	98	373452	41.211	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.420%
43) trans-1,3-Dichloroprop...	7.622	79	54820	39.588	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.180%
47) 2-Hexanone-d5	8.088	63	96165	79.498	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.500%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	169136	44.745	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.500%
66) 1,2-Dichlorobenzene-d4	11.622	152	127501	44.404	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024938.D
Acq On : 11 Mar 2022 00:53
Operator : SY/MD
Sample : N1879-07
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D97

Quant Time: Mar 11 05:27:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

