

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024945.D
 Acq On : 11 Mar 2022 03:39
 Operator : SY/MD
 Sample : N1856-15
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D72

Quant Time: Mar 11 05:29:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	281275	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	273204	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	122325	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	111977	39.152	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	78.300%	
7) Chloroethane-d5	1.571	69	88888	39.613	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	79.220%	
11) 1,1-Dichloroethene-d2	2.108	63	150289	31.308	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	62.620%	
21) 2-Butanone-d5	3.895	46	145695	89.513	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	89.510%	
24) Chloroform-d	4.349	84	203724	40.805	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	81.600%	
26) 1,2-Dichloroethane-d4	5.030	65	124184	42.157	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.320%	
32) Benzene-d6	5.050	84	429702	42.434	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.860%	
36) 1,2-Dichloropropane-d6	6.069	67	134948	42.773	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	85.540%	
41) Toluene-d8	7.313	98	371725	41.435	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	82.880%	
43) trans-1,3-Dichloroprop...	7.622	79	53531	39.048	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.100%	
47) 2-Hexanone-d5	8.088	63	95405	79.666	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	79.670%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	164446	43.944	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	87.880%	
66) 1,2-Dichlorobenzene-d4	11.622	152	125625	44.583	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.160%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.121	96	10750	5.163	ug/L	# 1
20) cis-1,2-Dichloroethene	3.915	96	9419	3.864	ug/L	98
30) 1,1,1-Trichloroethane	4.609	97	5359	1.490	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024945.D
Acq On : 11 Mar 2022 03:39
Operator : SY/MD
Sample : N1856-15
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D72

Quant Time: Mar 11 05:29:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

