

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024947.D
 Acq On : 11 Mar 2022 04:26
 Operator : SY/MD
 Sample : N1856-17
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D75

Quant Time: Mar 11 05:29:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	278128	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	272334	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	120661	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	109616	38.760	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	77.520%	
7) Chloroethane-d5	1.568	69	89426	40.304	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	80.600%	
11) 1,1-Dichloroethene-d2	2.108	63	146321	30.826	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.660%	
21) 2-Butanone-d5	3.896	46	144783	89.959	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	89.960%	
24) Chloroform-d	4.349	84	205189	41.563	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.120%	
26) 1,2-Dichloroethane-d4	5.031	65	124617	42.783	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.560%	
32) Benzene-d6	5.050	84	428808	42.481	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.960%	
36) 1,2-Dichloropropane-d6	6.069	67	134967	42.916	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	85.840%	
41) Toluene-d8	7.313	98	371643	41.559	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	83.120%	
43) trans-1,3-Dichloroprop...	7.622	79	53612	39.232	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.460%	
47) 2-Hexanone-d5	8.088	63	96088	80.493	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	80.490%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	164770	44.172	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.340%	
66) 1,2-Dichlorobenzene-d4	11.622	152	125312	45.085	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.180%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024947.D
Acq On : 11 Mar 2022 04:26
Operator : SY/MD
Sample : N1856-17
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D75

Quant Time: Mar 11 05:29:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

