

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024949.D
 Acq On : 11 Mar 2022 05:14
 Operator : SY/MD
 Sample : N1856-19
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D78

Quant Time: Mar 11 07:25:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	277128	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	272984	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	120303	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	110602	39.250	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.500%
7) Chloroethane-d5	1.571	69	87253	39.467	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.940%
11) 1,1-Dichloroethene-d2	2.108	63	144854	30.627	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.260%
21) 2-Butanone-d5	3.896	46	133872	83.480	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.480%
24) Chloroform-d	4.349	84	200226	40.704	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.400%
26) 1,2-Dichloroethane-d4	5.031	65	121820	41.973	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.940%
32) Benzene-d6	5.050	84	427474	42.248	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.500%
36) 1,2-Dichloropropane-d6	6.069	67	132881	42.152	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.300%
41) Toluene-d8	7.317	98	369545	41.226	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.460%
43) trans-1,3-Dichloroprop...	7.622	79	53190	38.831	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.660%
47) 2-Hexanone-d5	8.088	63	89706	74.968	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.970%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	161068	43.076	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.160%
66) 1,2-Dichlorobenzene-d4	11.622	152	124198	44.817	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024949.D
Acq On : 11 Mar 2022 05:14
Operator : SY/MD
Sample : N1856-19
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D78

Quant Time: Mar 11 07:25:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

