

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031022\
 Data File : VW024913.D
 Acq On : 10 Mar 2022 14:10
 Operator : SY/MD
 Sample : N1855-14
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D41

Quant Time: Mar 11 02:47:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 02:45:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	292301	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	285171	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	128053	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	131529	44.254	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.500%
7) Chloroethane-d5	1.568	69	101303	43.443	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.880%
11) 1,1-Dichloroethene-d2	2.108	63	165335	33.143	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.280%
21) 2-Butanone-d5	3.895	46	139654	82.565	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.560%
24) Chloroform-d	4.349	84	212353	40.929	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.860%
26) 1,2-Dichloroethane-d4	5.030	65	128227	41.887	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.780%
32) Benzene-d6	5.050	84	452414	42.802	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.069	67	140632	42.704	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.400%
41) Toluene-d8	7.313	98	394642	42.144	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.280%
43) trans-1,3-Dichloroprop...	7.622	79	58114	40.612	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.220%
47) 2-Hexanone-d5	8.088	63	90851	72.680	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.680%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	162459	41.592	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.180%
66) 1,2-Dichlorobenzene-d4	11.622	152	131825	44.690	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024913.D
Acq On : 10 Mar 2022 14:10
Operator : SY/MD
Sample : N1855-14
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D41

Quant Time: Mar 11 02:47:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 02:45:24 2022
Response via : Initial Calibration

