

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024933.D
 Acq On : 10 Mar 2022 22:55
 Operator : SY/MD
 Sample : N1879-02
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D85

Quant Time: Mar 11 05:26:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	290978	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	281552	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	126412	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	116110	39.244	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	78.480%	
7) Chloroethane-d5	1.568	69	92831	39.991	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	79.980%	
11) 1,1-Dichloroethene-d2	2.108	63	152226	30.654	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.300%	
21) 2-Butanone-d5	3.892	46	143103	84.988	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	84.990%	
24) Chloroform-d	4.349	84	206758	40.032	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	80.060%	
26) 1,2-Dichloroethane-d4	5.031	65	126350	41.462	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.920%	
32) Benzene-d6	5.050	84	435730	41.753	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.500%	
36) 1,2-Dichloropropane-d6	6.069	67	138868	42.710	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	85.420%	
41) Toluene-d8	7.313	98	383786	41.511	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	83.020%	
43) trans-1,3-Dichloroprop...	7.622	79	56217	39.791	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	79.580%	
47) 2-Hexanone-d5	8.088	63	96026	77.807	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	77.810%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	170485	44.207	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.420%	
66) 1,2-Dichlorobenzene-d4	11.622	152	128433	44.106	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.220%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024933.D
Acq On : 10 Mar 2022 22:55
Operator : SY/MD
Sample : N1879-02
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D85

Quant Time: Mar 11 05:26:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

