

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024935.D
 Acq On : 10 Mar 2022 23:42
 Operator : SY/MD
 Sample : N1879-04
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D88

Quant Time: Mar 11 05:26:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	286344	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	275288	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	125197	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	116759	40.102	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.200%
7) Chloroethane-d5	1.568	69	92428	40.462	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.920%
11) 1,1-Dichloroethene-d2	2.108	63	150728	30.843	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.680%
21) 2-Butanone-d5	3.896	46	140520	84.805	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.800%
24) Chloroform-d	4.349	84	208838	41.089	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.180%
26) 1,2-Dichloroethane-d4	5.031	65	126879	42.309	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.620%
32) Benzene-d6	5.050	84	441814	43.299	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	6.069	67	138477	43.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.120%
41) Toluene-d8	7.313	98	381078	42.156	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.320%
43) trans-1,3-Dichloroprop...	7.619	79	56996	41.261	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.520%
47) 2-Hexanone-d5	8.088	63	95479	79.125	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.120%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	167844	44.513	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.020%
66) 1,2-Dichlorobenzene-d4	11.622	152	129366	44.857	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024935.D
Acq On : 10 Mar 2022 23:42
Operator : SY/MD
Sample : N1879-04
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D88

Quant Time: Mar 11 05:26:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

