

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
 Data File : VV024941.D
 Acq On : 11 Mar 2022 02:04
 Operator : SY/MD
 Sample : N1856-11
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D66

Quant Time: Mar 11 05:28:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 05:21:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	285750	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	282497	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	124150	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	113837	39.179	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	78.360%	
7) Chloroethane-d5	1.571	69	88999	39.042	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	78.080%	
11) 1,1-Dichloroethene-d2	2.108	63	148823	30.517	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.040%	
21) 2-Butanone-d5	3.895	46	141766	85.735	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	85.730%	
24) Chloroform-d	4.349	84	205886	40.592	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	81.180%	
26) 1,2-Dichloroethane-d4	5.030	65	125625	41.978	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.960%	
32) Benzene-d6	5.050	84	430789	41.142	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.280%	
36) 1,2-Dichloropropane-d6	6.069	67	136659	41.890	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	83.780%	
41) Toluene-d8	7.317	98	375307	40.458	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	80.920%	
43) trans-1,3-Dichloroprop...	7.622	79	53411	37.679	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	75.360%	
47) 2-Hexanone-d5	8.088	63	97369	78.632	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	78.630%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	167672	43.332	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	86.660%	
66) 1,2-Dichlorobenzene-d4	11.622	152	126189	44.125	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031022\
Data File : VV024941.D
Acq On : 11 Mar 2022 02:04
Operator : SY/MD
Sample : N1856-11
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D66

Quant Time: Mar 11 05:28:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 05:21:24 2022
Response via : Initial Calibration

