

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031122\
 Data File : VW024961.D
 Acq On : 11 Mar 2022 13:47
 Operator : SY/MD
 Sample : N1879-15
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 11 23:10:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 23:08:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	275219	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	269697	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	119196	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	147863	52.837	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.680%
7) Chloroethane-d5	1.568	69	111746	50.896	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.800%
11) 1,1-Dichloroethene-d2	2.108	63	182486	38.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.700%
21) 2-Butanone-d5	3.896	46	157019	98.593	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.590%
24) Chloroform-d	4.349	84	238010	48.721	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.440%
26) 1,2-Dichloroethane-d4	5.031	65	145273	50.401	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.800%
32) Benzene-d6	5.050	84	512715	51.290	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.580%
36) 1,2-Dichloropropane-d6	6.069	67	156341	50.198	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.400%
41) Toluene-d8	7.313	98	446618	50.431	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.860%
43) trans-1,3-Dichloroprop...	7.619	79	64499	47.660	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.320%
47) 2-Hexanone-d5	8.088	63	105589	89.317	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.320%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	182888	49.508	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.020%
66) 1,2-Dichlorobenzene-d4	11.622	152	149271	54.365	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031122\
Data File : VV024961.D
Acq On : 11 Mar 2022 13:47
Operator : SY/MD
Sample : N1879-15
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 11 23:10:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 23:08:36 2022
Response via : Initial Calibration

