

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031122\
 Data File : VW024967.D
 Acq On : 11 Mar 2022 16:33
 Operator : SY/MD
 Sample : N1880-05
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DB5

Quant Time: Mar 11 23:11:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 11 23:08:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	313233	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	303297	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	138192	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	144384	45.333	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.660%
7) Chloroethane-d5	1.568	69	111062	44.445	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.900%
11) 1,1-Dichloroethene-d2	2.108	63	184283	34.473	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.940%
21) 2-Butanone-d5	3.892	46	156434	86.305	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.300%
24) Chloroform-d	4.349	84	243028	43.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.420%
26) 1,2-Dichloroethane-d4	5.030	65	146017	44.511	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.020%
32) Benzene-d6	5.050	84	519988	46.255	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.500%
36) 1,2-Dichloropropane-d6	6.069	67	160134	45.720	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.440%
41) Toluene-d8	7.313	98	454019	45.587	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.180%
43) trans-1,3-Dichloroprop...	7.619	79	68214	44.821	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.640%
47) 2-Hexanone-d5	8.088	63	108370	81.514	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.510%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	186563	44.908	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.820%
66) 1,2-Dichlorobenzene-d4	11.622	152	151574	47.616	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031122\
Data File : VV024967.D
Acq On : 11 Mar 2022 16:33
Operator : SY/MD
Sample : N1880-05
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DB5

Quant Time: Mar 11 23:11:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 11 23:08:36 2022
Response via : Initial Calibration

