

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031422\
 Data File : VW024980.D
 Acq On : 14 Mar 2022 11:51
 Operator : SY/MD
 Sample : N1880-13
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DC7

Quant Time: Mar 15 02:20:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 15 02:18:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	311827	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	304632	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	136540	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	148240	46.753	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.500%
7) Chloroethane-d5	1.571	69	116565	46.858	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.720%
11) 1,1-Dichloroethene-d2	2.111	63	188312	35.385	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.780%
21) 2-Butanone-d5	3.895	46	160258	88.813	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.810%
24) Chloroform-d	4.349	84	247550	44.725	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.460%
26) 1,2-Dichloroethane-d4	5.030	65	148683	45.528	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.060%
32) Benzene-d6	5.050	84	528164	46.776	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.560%
36) 1,2-Dichloropropane-d6	6.069	67	165034	46.912	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.820%
41) Toluene-d8	7.313	98	454127	45.398	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.622	79	64039	41.894	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.780%
47) 2-Hexanone-d5	8.088	63	110693	82.896	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.900%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	196036	46.982	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.960%
66) 1,2-Dichlorobenzene-d4	11.622	152	152861	48.601	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031422\
Data File : VV024980.D
Acq On : 14 Mar 2022 11:51
Operator : SY/MD
Sample : N1880-13
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DC7

Quant Time: Mar 15 02:20:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 15 02:18:50 2022
Response via : Initial Calibration

