

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031422\
 Data File : VW024982.D
 Acq On : 14 Mar 2022 13:03
 Operator : SY/MD
 Sample : N1880-16
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CODE0

Quant Time: Mar 15 02:20:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 15 02:18:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	299236	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	292321	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	130758	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	123510	40.593	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.180%
7) Chloroethane-d5	1.568	69	101266	42.421	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.840%
11) 1,1-Dichloroethene-d2	2.108	63	160727	31.473	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.940%
21) 2-Butanone-d5	3.892	46	142506	82.298	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.300%
24) Chloroform-d	4.346	84	220651	41.543	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.080%
26) 1,2-Dichloroethane-d4	5.030	65	135071	43.100	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.200%
32) Benzene-d6	5.050	84	462386	42.675	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.360%
36) 1,2-Dichloropropane-d6	6.069	67	147477	43.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.380%
41) Toluene-d8	7.313	98	399006	41.568	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.140%
43) trans-1,3-Dichloroprop...	7.619	79	62535	42.633	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.260%
47) 2-Hexanone-d5	8.088	63	98519	76.887	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.890%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	174929	43.689	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.380%
66) 1,2-Dichlorobenzene-d4	11.622	152	136175	45.210	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031422\
Data File : VV024982.D
Acq On : 14 Mar 2022 13:03
Operator : SY/MD
Sample : N1880-16
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CODE0

Quant Time: Mar 15 02:20:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 15 02:18:50 2022
Response via : Initial Calibration

