

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031422\
 Data File : VW024984.D
 Acq On : 14 Mar 2022 13:51
 Operator : SY/MD
 Sample : N1880-18
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CODE5

Quant Time: Mar 15 02:21:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 15 02:18:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	319571	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	306886	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	139791	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	132698	40.837	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.680%
7) Chloroethane-d5	1.571	69	105096	41.224	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.440%
11) 1,1-Dichloroethene-d2	2.108	63	170975	31.349	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.700%
21) 2-Butanone-d5	3.896	46	164749	89.089	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.090%
24) Chloroform-d	4.349	84	234633	41.364	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.720%
26) 1,2-Dichloroethane-d4	5.031	65	142409	42.550	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.100%
32) Benzene-d6	5.050	84	493926	43.422	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.840%
36) 1,2-Dichloropropane-d6	6.069	67	157084	44.324	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.640%
41) Toluene-d8	7.314	98	425156	42.190	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.380%
43) trans-1,3-Dichloroprop...	7.619	79	67012	43.517	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.040%
47) 2-Hexanone-d5	8.088	63	114471	85.096	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.100%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	191698	45.604	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	146894	45.618	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031422\
Data File : VV024984.D
Acq On : 14 Mar 2022 13:51
Operator : SY/MD
Sample : N1880-18
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DE5

Quant Time: Mar 15 02:21:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 15 02:18:50 2022
Response via : Initial Calibration

