

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031424\
 Data File : VW034592.D
 Acq On : 14 Mar 2024 10:40
 Operator : SY/MD
 Sample : P1739-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GE1

Quant Time: Mar 15 03:34:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 03:33:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	307651	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	288143	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	133676	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	134141	50.296	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.600%		
7) Chloroethane-d5	1.529	69	111147	53.493	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 106.980%		
11) 1,1-Dichloroethene-d2	2.056	65	61538	51.020	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 102.040%		
21) 2-Butanone-d5	3.786	46	172983	96.989	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 96.990%		
24) Chloroform-d	4.243	84	242700	51.468	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.940%		
26) 1,2-Dichloroethane-d4	4.937	65	163417	53.032	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.060%		
32) Benzene-d6	4.957	84	476322	54.773	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 109.540%		
36) 1,2-Dichloropropane-d6	5.985	67	145262	54.735	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 109.460%		
41) Toluene-d8	7.239	98	418173	53.137	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.280%		
43) trans-1,3-Dichloroprop...	7.551	79	65415	49.084	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 98.160%		
47) 2-Hexanone-d5	8.021	63	127111	95.915	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 95.910%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	178876	51.217	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 102.440%		
66) 1,2-Dichlorobenzene-d4	11.558	152	149200	54.604	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
Data File : VV034592.D
Acq On : 14 Mar 2024 10:40
Operator : SY/MD
Sample : P1739-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GE1

Quant Time: Mar 15 03:34:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 03:33:21 2024
Response via : Initial Calibration

