

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031424\
 Data File : VW034602.D
 Acq On : 14 Mar 2024 14:39
 Operator : SY/MD
 Sample : P1740-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GH1

Quant Time: Mar 15 03:36:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 03:33:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	299751	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	281907	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	135510	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	126243	48.583	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.160%
7) Chloroethane-d5	1.532	69	105072	51.902	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.800%
11) 1,1-Dichloroethene-d2	2.059	65	56730	48.273	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.540%
21) 2-Butanone-d5	3.796	46	174172	100.229	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.230%
24) Chloroform-d	4.249	84	223656	48.679	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.360%
26) 1,2-Dichloroethane-d4	4.940	65	150235	50.039	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.080%
32) Benzene-d6	4.960	84	434399	51.058	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.120%
36) 1,2-Dichloropropane-d6	5.989	67	135314	52.114	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.220%
41) Toluene-d8	7.243	98	386917	50.253	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.500%
43) trans-1,3-Dichloroprop...	7.554	79	59081	45.312	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.620%
47) 2-Hexanone-d5	8.024	63	122641	94.589	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.590%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	171815	50.283	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.560%
66) 1,2-Dichlorobenzene-d4	11.561	152	143527	51.817	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
Data File : VV034602.D
Acq On : 14 Mar 2024 14:39
Operator : SY/MD
Sample : P1740-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GH1

Quant Time: Mar 15 03:36:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 03:33:21 2024
Response via : Initial Calibration

