

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
 Data File : VV034614.D
 Acq On : 14 Mar 2024 19:26
 Operator : SY/MD
 Sample : P1751-22
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL55

Quant Time: Mar 15 03:39:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 03:33:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	284025	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	262675	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	124204	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	106829	43.388	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.780%	
7) Chloroethane-d5	1.532	69	91349	47.622	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	95.240%	
11) 1,1-Dichloroethene-d2	2.060	65	49118	44.110	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.220%	
21) 2-Butanone-d5	3.796	46	173838	105.576	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	=	105.580%	
24) Chloroform-d	4.246	84	207487	47.661	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.320%	
26) 1,2-Dichloroethane-d4	4.941	65	141344	49.684	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.360%	
32) Benzene-d6	4.957	84	397006	50.079	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.160%	
36) 1,2-Dichloropropane-d6	5.985	67	125810	52.002	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.000%	
41) Toluene-d8	7.243	98	342461	47.736	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.480%	
43) trans-1,3-Dichloroprop...	7.555	79	56403	46.425	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	92.860%	
47) 2-Hexanone-d5	8.024	63	119089	98.574	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	98.570%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	161482	50.719	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.440%	
66) 1,2-Dichlorobenzene-d4	11.561	152	130855	51.543	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.080%	
Target Compounds						
13) Acetone	2.130	43	15503	7.832	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
Data File : VV034614.D
Acq On : 14 Mar 2024 19:26
Operator : SY/MD
Sample : P1751-22
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCL55

Quant Time: Mar 15 03:39:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 03:33:21 2024
Response via : Initial Calibration

