

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
 Data File : VV034618.D
 Acq On : 14 Mar 2024 21:26
 Operator : SY/MD
 Sample : P1751-14
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL19

Quant Time: Mar 15 06:54:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 06:40:01 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	238864	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	223678	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	97120	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	112510	54.334	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery	= 108.660%		
7) Chloroethane-d5	1.532	69	99241	61.517	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery	= 123.040%		
11) 1,1-Dichloroethene-d2	2.059	65	52432	55.989	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery	= 111.980%		
21) 2-Butanone-d5	3.796	46	190517	137.581	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery	= 137.580##		
24) Chloroform-d	4.249	84	215160	58.767	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery	= 117.540%		
26) 1,2-Dichloroethane-d4	4.940	65	147353	61.589	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery	= 123.180%		
32) Benzene-d6	4.960	84	421236	62.399	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery	= 124.800%		
36) 1,2-Dichloropropane-d6	5.989	67	131269	63.718	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery	= 127.440##		
41) Toluene-d8	7.243	98	363730	59.540	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery	= 119.080%		
43) trans-1,3-Dichloroprop...	7.554	79	55472	53.619	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery	= 107.240%		
47) 2-Hexanone-d5	8.024	63	128540	124.947	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery	= 124.950%		
56) 1,1,2,2-Tetrachloroeth...	10.152	84	175261	64.644	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery	= 129.280##		
66) 1,2-Dichlorobenzene-d4	11.561	152	138090	69.561	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery	= 139.120##		
Target Compounds						
5) Vinyl chloride	1.281	62	4871	2.038	ug/L	# 43
12) 1,1-Dichloroethene	2.069	96	70163	39.228	ug/L	82
13) Acetone	2.127	43	337333	202.638	ug/L	99
19) 1,1-Dichloroethane	3.114	63	15352	4.527	ug/L	97
20) cis-1,2-Dichloroethene	3.815	96	20762	11.053	ug/L	94
30) 1,1,1-Trichloroethane	4.516	97	14007	4.367	ug/L	98
34) Trichloroethene	5.841	95	22889	11.598	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
Data File : VV034618.D
Acq On : 14 Mar 2024 21:26
Operator : SY/MD
Sample : P1751-14
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCL19

Quant Time: Mar 15 06:54:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 06:40:01 2024
Response via : Initial Calibration

