

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
 Data File : VV034625.D
 Acq On : 15 Mar 2024 00:37
 Operator : SY/MD
 Sample : P1751-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCKX6

Quant Time: Mar 15 06:56:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 06:40:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	264925	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	250845	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	115552	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	93295	40.623	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.240%
7) Chloroethane-d5	1.526	69	80560	45.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.040%
11) 1,1-Dichloroethene-d2	2.056	65	43407	41.792	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.580%
21) 2-Butanone-d5	3.789	46	172119	112.068	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.070%
24) Chloroform-d	4.246	84	198440	48.869	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.740%
26) 1,2-Dichloroethane-d4	4.940	65	141398	53.286	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.580%
32) Benzene-d6	4.956	84	380933	50.318	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.640%
36) 1,2-Dichloropropane-d6	5.989	67	121396	52.543	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.080%
41) Toluene-d8	7.243	98	325859	47.564	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.120%
43) trans-1,3-Dichloroprop...	7.554	79	53384	46.013	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.020%
47) 2-Hexanone-d5	8.024	63	117955	102.240	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.240%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	168674	55.477	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.960%
66) 1,2-Dichlorobenzene-d4	11.561	152	122431	51.835	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.680%
Target Compounds						
13) Acetone	2.117	43	401703	217.568	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031424\
Data File : VV034625.D
Acq On : 15 Mar 2024 00:37
Operator : SY/MD
Sample : P1751-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCKX6

Quant Time: Mar 15 06:56:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 06:40:01 2024
Response via : Initial Calibration

