

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV031425\
Data File : VV038708.D
Acq On : 14 Mar 2025 10:09
Operator : SY/MD
Sample : VV0314WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 14 11:06:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031325WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 13 13:15:04 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	146281	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	158548	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	61103	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	44095	4.55	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.00%	
7) Chloroethane-d5	1.529	69	43624	4.63	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.60%	
11) 1,1-Dichloroethene-d2	2.053	65	22588	4.57	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.40%	
20) 2-Butanone-d5	3.802	46	101260	43.40	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.80%	
24) Chloroform-d	4.240	84	107617	4.72	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.40%	
26) 1,2-Dichloroethane-d4	4.937	65	51197	5.01	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.20%	
32) Benzene-d6	4.953	84	189330	4.65	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.00%	
36) 1,2-Dichloropropane-d6	5.985	67	63635	4.81	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.20%	
41) Toluene-d8	7.239	98	142652	3.88	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.60%	
43) trans-1,3-Dichloroprop...	7.561	79	15464	3.97	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.40%	
46) 2-Hexanone-d5	8.021	63	77026	40.99	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	81.98%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	52091	4.66	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.20%	
66) 1,2-Dichlorobenzene-d4	11.554	152	52218	5.24	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.80%	

Target Compounds

ABTKS84
Client Sample :
MSVOA_V
Instrument :

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV031425\
 Data File : VV038708.D
 Acq On : 14 Mar 2025 10:09
 Operator : SY/MD
 Sample : VV0314WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 14 11:06:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031325WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 13 13:15:04 2025
 Response via : Initial Calibration

