

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031522\
 Data File : VW025005.D
 Acq On : 15 Mar 2022 18:33
 Operator : SY/MD
 Sample : N1934-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDD7

Quant Time: Mar 16 03:40:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 16 03:30:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	166249	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	155679	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68974	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	57750	4.400	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.000%
7) Chloroethane-d5	1.571	69	45636	4.748	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.111	63	72455	3.416	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.400%
20) 2-Butanone-d5	3.892	46	93351	58.851	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.700%
24) Chloroform-d	4.352	84	101233	4.650	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.034	65	44084	4.844	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	5.053	84	208922	4.626	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.069	67	62426	4.807	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.317	98	181498	4.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	7.625	79	19962	4.409	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.200%
46) 2-Hexanone-d5	8.088	63	74640	48.336	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.680%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	36574	4.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	58387	5.000	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds						
					Qvalue	
13) Acetone	2.185	43	24432	23.030	ug/L	84
21) 2-Butanone	4.011	43	3820	2.281	ug/L #	78
31) Carbon tetrachloride	4.831	117	17159	1.007	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031522\
Data File : VV025005.D
Acq On : 15 Mar 2022 18:33
Operator : SY/MD
Sample : N1934-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDD7

Quant Time: Mar 16 03:40:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 16 03:30:21 2022
Response via : Initial Calibration

