

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031522\
 Data File : VW025010.D
 Acq On : 15 Mar 2022 20:30
 Operator : SY/MD
 Sample : N1933-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDB7

Quant Time: Mar 16 03:41:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 16 03:30:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	163459	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	152625	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67836	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62186	4.819	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.400%	
7) Chloroethane-d5	1.571	69	48591	5.142	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.800%	
11) 1,1-Dichloroethene-d2	2.111	63	79572	3.815	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.400%	
20) 2-Butanone-d5	3.892	46	93757	60.116	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.240%	
24) Chloroform-d	4.352	84	104449	4.880	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
26) 1,2-Dichloroethane-d4	5.037	65	45457	5.080	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.600%	
32) Benzene-d6	5.053	84	221627	5.006	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.200%	
36) 1,2-Dichloropropane-d6	6.069	67	65363	5.134	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.600%	
41) Toluene-d8	7.317	98	193596	4.798	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.000%	
43) trans-1,3-Dichloroprop...	7.625	79	20176	4.546	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.000%	
46) 2-Hexanone-d5	8.088	63	71696	47.359	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.720%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	38580	5.151	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	60762	5.291	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.800%	
Target Compounds						
					Qvalue	
13) Acetone	2.178	43	7844	7.520	ug/L	79
25) Chloroform	4.378	83	2957	0.133	ug/L	83
31) Carbon tetrachloride	4.831	117	137863	8.256	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031522\
Data File : VV025010.D
Acq On : 15 Mar 2022 20:30
Operator : SY/MD
Sample : N1933-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDB7

Quant Time: Mar 16 03:41:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 16 03:30:21 2022
Response via : Initial Calibration

