

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031622\
 Data File : VW025044.D
 Acq On : 16 Mar 2022 13:52
 Operator : SY/MD
 Sample : N1933-15DL 50X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDC6DL

Quant Time: Mar 17 02:02:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 17 01:58:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	155875	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	146957	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63772	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	52143	4.237	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.800%	
7) Chloroethane-d5	1.568	69	42121	4.674	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.400%	
11) 1,1-Dichloroethene-d2	2.108	63	66155	3.326	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.600%	
20) 2-Butanone-d5	3.912	46	83207	55.947	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.900%	
24) Chloroform-d	4.352	84	94707	4.640	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	5.034	65	41511	4.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.050	84	195480	4.585	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.800%	
36) 1,2-Dichloropropane-d6	6.072	67	58872	4.802	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.000%	
41) Toluene-d8	7.317	98	166811	4.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.800%	
43) trans-1,3-Dichloroprop...	7.625	79	17333	4.056	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.200%	
46) 2-Hexanone-d5	8.092	63	66979	45.949	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.900%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34585	4.796	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	53888	4.992	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.800%	
Target Compounds						
					Qvalue	
25) Chloroform	4.381	83	2602	0.123	ug/L	92
31) Carbon tetrachloride	4.831	117	94707	5.890	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025044.D
Acq On : 16 Mar 2022 13:52
Operator : SY/MD
Sample : N1933-15DL 50X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDC6DL

Quant Time: Mar 17 02:02:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 17 01:58:56 2022
Response via : Initial Calibration

