

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031622\
 Data File : VW025053.D
 Acq On : 16 Mar 2022 17:23
 Operator : SY/MD
 Sample : N1941-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JNLG5

Quant Time: Mar 17 02:04:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 17 01:58:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	150011	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	144206	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	48849	4.125	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.400%
7) Chloroethane-d5	1.571	69	40596	4.681	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.600%
11) 1,1-Dichloroethene-d2	2.111	63	62758	3.279	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.905	46	86262	60.269	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.540%
24) Chloroform-d	4.352	84	91324	4.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.037	65	41249	5.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.053	84	187697	4.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.072	67	56758	4.718	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.317	98	159852	4.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
43) trans-1,3-Dichloroprop...	7.625	79	18258	4.354	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.091	63	63152	44.150	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.300%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31156	4.403	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	51347	5.034	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%
Target Compounds						Qvalue
3) Chloromethane	1.243	50	2550	0.204	ug/L	99
25) Chloroform	4.381	83	6788	0.333	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025053.D
Acq On : 16 Mar 2022 17:23
Operator : SY/MD
Sample : N1941-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JNLG5

Quant Time: Mar 17 02:04:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 17 01:58:56 2022
Response via : Initial Calibration

