

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031622\
 Data File : VW025052.D
 Acq On : 16 Mar 2022 17:00
 Operator : SY/MD
 Sample : N1933-21DL 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDD2DL

Quant Time: Mar 17 02:04:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 17 01:58:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151685	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	143220	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61817	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	48202	4.025	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.600%
7) Chloroethane-d5	1.571	69	40305	4.596	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.111	63	63116	3.261	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	3.896	46	88929	61.447	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.900%
24) Chloroform-d	4.352	84	89651	4.514	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.034	65	40017	4.819	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.050	84	183772	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.069	67	56700	4.746	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.317	98	156916	4.144	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	7.625	79	17365	4.169	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.088	63	65176	45.879	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.760%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33233	4.728	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	50327	4.809	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%
Target Compounds						
					Qvalue	
13) Acetone	2.188	43	23742	24.529	ug/L	83
16) Methylene chloride	2.510	84	2742	0.234	ug/L	97
21) 2-Butanone	3.992	43	16533	10.822	ug/L	79
25) Chloroform	4.384	83	2323	0.113	ug/L	95
31) Carbon tetrachloride	4.831	117	63391	4.045	ug/L	98
42) Toluene	7.397	91	11455	0.235	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025052.D
Acq On : 16 Mar 2022 17:00
Operator : SY/MD
Sample : N1933-21DL 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDD2DL

Quant Time: Mar 17 02:04:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 17 01:58:56 2022
Response via : Initial Calibration





#13

Acetone

Concen: 24.529 ug/L

RT: 2.188 min Scan# 31

Delta R.T. 0.000 min

Lab File: VV025052.D

Acq: 16 Mar 2022 17:00

Instrument :

MSVOA_V

ClientSampleId :





GBDD2DL

Tgt Ion: 43 Resp: 23742

Ion Ratio Lower Upper

43 100

58 36.2 0.0 54.2



#16

Methylene chloride

Concen: 0.234 ug/L

RT: 2.510 min Scan# 457

Delta R.T. 0.003 min

Lab File: VV025052.D

Acq: 16 Mar 2022 17:00




Tgt Ion: 84 Resp: 2742


Ion Ratio Lower Upper

84 100

86 68.8 45.5 84.5

49 109.4 75.0 139.2









#21
2-Butanone
Concen: 10.822 ug/L
RT: 3.992 min Scan# 913
Delta R.T. 0.016 min
Lab File: VV025052.D
Acq: 16 Mar 2022 17:00




Instrument :
MSVOA_V
ClientSampleId :
GBDD2DL


Tgt Ion: 43 Resp: 16533
Ion Ratio Lower Upper
43 100
72 30.5 10.4 31.2



#25
Chloroform
Concen: 0.113 ug/L
RT: 4.384 min Scan# 1040
Delta R.T. 0.013 min
Lab File: VV025052.D
Acq: 16 Mar 2022 17:00

Tgt Ion: 83 Resp: 2323
Ion Ratio Lower Upper
83 100
85 59.7 44.5 82.7









#31
Carbon tetrachloride
Concen: 4.045 ug/L
RT: 4.831 min Scan# 1177
Delta R.T. 0.007 min
Lab File: VV025052.D
Acq: 16 Mar 2022 17:00

Instrument :
MSVOA_V
ClientSampleId :
GBDD2DL

Tgt Ion:117 Resp: 63391
Ion Ratio Lower Upper
117 100
119 94.0 76.8 115.2



#42
Toluene
Concen: 0.235 ug/L
RT: 7.397 min Scan# 1977
Delta R.T. 0.013 min
Lab File: VV025052.D
Acq: 16 Mar 2022 17:00

Tgt Ion: 91 Resp: 11455
Ion Ratio Lower Upper
91 100
92 58.4 41.2 76.4

