

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW031622\  
 Data File : VW025070.D  
 Acq On : 17 Mar 2022 00:49  
 Operator : SY/MD  
 Sample : N1941-14  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 39 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 JNLG4

Quant Time: Mar 17 06:29:21 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR031522WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Mar 17 05:44:36 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 5.619  | 114            | 151540   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117            | 144654   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 63060    | 5.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.307  | 65             | 44843    | 3.748  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 75.000%  |          |
| 7) Chloroethane-d5            | 1.571  | 69             | 36579    | 4.175  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 83.600%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63             | 57663    | 2.982  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 59.600%# |          |
| 20) 2-Butanone-d5             | 3.899  | 46             | 91895    | 63.557 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 127.120% |          |
| 24) Chloroform-d              | 4.352  | 84             | 88450    | 4.458  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 89.200%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.037  | 65             | 40033    | 4.826  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 96.600%  |          |
| 32) Benzene-d6                | 5.053  | 84             | 175138   | 4.174  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 83.400%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.072  | 67             | 54179    | 4.490  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 89.800%  |          |
| 41) Toluene-d8                | 7.317  | 98             | 146935   | 3.842  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 76.800%  |          |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79             | 16279    | 3.870  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 77.400%  |          |
| 46) 2-Hexanone-d5             | 8.088  | 63             | 70151    | 48.892 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 97.780%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214 | 84             | 35143    | 4.951  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 99.000%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152            | 50574    | 4.737  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 94.800%  |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV031622\  
Data File : VV025070.D  
Acq On : 17 Mar 2022 00:49  
Operator : SY/MD  
Sample : N1941-14  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 39 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
JNLG4

Quant Time: Mar 17 06:29:21 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR031522WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Mar 17 05:44:36 2022  
Response via : Initial Calibration

