

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
 Data File : VV025049.D
 Acq On : 16 Mar 2022 15:50
 Operator : SY/MD
 Sample : N1933-19
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDD0

Integration Parameters: rteint.p

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.1

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Title : TRACE VOA SFAM1.0

Signal : TIC: VV025049.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.108	16	21	27	rBV	26441	24592	5.03%	0.575%
2	1.307	76	83	97	rVB	79357	81825	16.75%	1.913%
3	1.571	157	165	176	rBV	63826	71443	14.63%	1.670%
4	2.111	322	333	345	rBV	127070	181158	37.09%	4.235%
5	2.182	346	355	383	rVB	19277	49359	10.10%	1.154%
6	2.355	393	409	411	rBV6	3889	8420	1.72%	0.197%
7	3.892	878	887	931	rBV	44248	156660	32.07%	3.662%
8	4.085	938	947	973	rVB2	10785	30379	6.22%	0.710%
9	4.349	1012	1029	1055	rBV4	124217	330185	67.59%	7.719%
10	4.831	1165	1179	1197	rVB2	9380	21644	4.43%	0.506%
11	5.050	1231	1247	1272	rBV2	199002	488486	100.00%	11.420%
12	5.619	1412	1424	1456	rBV	178297	360891	73.88%	8.437%
13	5.912	1503	1515	1529	rVB6	5247	11333	2.32%	0.265%
14	6.069	1551	1564	1589	rBV	112093	228006	46.68%	5.330%
15	6.989	1839	1850	1879	rBV	49587	93998	19.24%	2.197%
16	7.317	1940	1952	1973	rBV	234706	410724	84.08%	9.602%
17	7.625	2032	2048	2067	rBV2	26859	49347	10.10%	1.154%
18	7.940	2135	2146	2157	rBV2	4596	7964	1.63%	0.186%
19	8.088	2183	2192	2209	rBV	178930	333263	68.22%	7.791%
20	8.850	2415	2429	2459	rBV	261756	433526	88.75%	10.135%
21	10.217	2842	2854	2875	rVB	66576	105445	21.59%	2.465%
22	10.381	2894	2905	2920	rVB2	9819	15548	3.18%	0.363%
23	11.249	3164	3175	3195	rBV	252889	387289	79.28%	9.054%
24	11.529	3252	3262	3276	rBV	32150	58027	11.88%	1.357%
25	11.622	3280	3291	3306	rVB	210191	332454	68.06%	7.772%
26	12.246	3477	3485	3501	rVB2	3509	5585	1.14%	0.131%


Sum of corrected areas: 4277551

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025049.D
Acq On : 16 Mar 2022 15:50
Operator : SY/MD
Sample : N1933-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDD0

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025049.D
Acq On : 16 Mar 2022 15:50
Operator : SY/MD
Sample : N1933-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDD0










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 1-Hexanol, 2-ethyl- Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
11.529	0.75 ug/L	58027	1,4-Dichlorobenzene-d4	11.249

Hit#	of 5	Tentative ID	MW	MolForm	CAS#	Qual
1		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	90
2		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	90
3		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	90
4		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	78
5		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	72



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
Data File : VV025049.D
Acq On : 16 Mar 2022 15:50
Operator : SY/MD
Sample : N1933-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDD0

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
1-Hexanol, 2-et...	11.529	0.8	ug/L	58027	3	11.249	387289	5.0