



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031622\
 Data File : VV025034.D
 Acq On : 16 Mar 2022 09:04
 Operator : SY/MD
 Sample : BFB280
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB280

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Mar 18 02:20:05 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	28232	PASS
75	95	30	80	49.8	88189	PASS
95	95	100	100	100.0	176960	PASS
96	95	5	9	6.7	11808	PASS
173	174	0.00	2	1.0	1212	PASS
174	95	50	120	71.9	127203	PASS
175	174	5	9	7.3	9303	PASS
176	174	95	105	95.1	121000	PASS
177	176	5	10	6.7	8133	PASS

m/z	Abundance
39.90	446.0
43.90	223.0

Instrument :
MSVOA_V
ClientSampleId :
BFB280

m/z	Abundance
35.90	1140.0
37.00	6711.0
38.00	5416.0
39.00	2479.0
39.90	628.0
44.00	1069.0
44.90	1207.0
45.90	113.0
47.00	2043.0
47.90	905.0
49.00	6095.0
50.00	28136.0
51.00	9451.0
52.00	427.0
54.90	426.0
55.90	2320.0
57.00	4677.0
57.90	167.0
59.90	1465.0
61.00	7539.0
62.00	7802.0
63.00	5913.0
64.00	873.0
65.00	815.0
66.90	431.0
68.00	17072.0
69.00	17904.0
70.00	1376.0
72.00	850.0
73.00	7516.0
74.00	28800.0
75.00	86624.0
76.00	7640.0
76.90	1023.0
77.90	785.0
78.90	5251.0
79.90	1480.0
80.90	5221.0
81.90	1237.0
82.80	150.0
85.90	269.0
86.90	7855.0
87.90	7837.0
90.80	690.0
92.00	4997.0
93.00	7209.0
94.00	20568.0
95.00	175552.0
96.00	11834.0
97.00	387.0
103.90	827.0
104.80	214.0
105.80	827.0
106.70	187.0
109.90	114.0
110.90	163.0
111.80	129.0
112.80	172.0
114.90	186.0
115.80	617.0
116.90	1167.0
117.80	545.0
118.80	853.0
123.90	116.0
127.90	661.0
128.80	343.0
129.90	596.0
130.80	210.0
134.90	337.0
136.80	300.0
139.70	103.0
140.90	1787.0
141.90	134.0
142.80	1841.0
143.90	169.0
144.80	315.0
145.90	190.0
147.80	340.0
148.80	117.0
149.90	195.0
152.70	108.0
153.70	106.0
154.90	460.0
156.80	299.0
158.80	164.0
160.80	219.0
170.80	158.0
171.80	962.0
173.00	1190.0
173.90	122344.0
174.90	9063.0
175.90	116480.0

Instrument :
MSVOA_V
ClientSampleId :
BFB280

176.907908.0

178.00263.0

Instrument :
MSVOA_V
ClientSampleId :
BFB280

m/z	Abundance
36.00	1192.0
37.00	6960.0
38.00	5591.0
39.00	2637.0
39.90	609.0
44.00	1104.0
44.90	1276.0
47.00	2136.0
47.90	955.0
49.00	6348.0
50.00	29104.0
51.00	9844.0
52.00	445.0
54.90	450.0
56.00	2354.0
57.00	4778.0
58.00	157.0
59.90	1577.0
61.00	7749.0
62.00	8226.0
63.00	6165.0
64.00	835.0
65.00	822.0
66.90	484.0
68.00	17640.0
69.00	18504.0
70.00	1418.0
72.00	913.0
73.00	7952.0
74.00	30320.0
75.00	91416.0
76.00	7841.0
76.90	1071.0
77.90	892.0
78.90	5572.0
79.90	1520.0
80.90	5415.0
81.90	1298.0
82.80	165.0
85.80	267.0
86.90	8082.0
87.90	8083.0
90.80	704.0
92.00	5337.0
93.00	7488.0
94.00	21680.0
95.00	183360.0
96.00	12269.0
97.00	371.0
103.90	864.0
104.80	249.0
105.80	874.0
106.70	190.0
109.70	107.0
110.90	166.0
111.80	139.0
112.80	129.0
114.90	203.0
115.80	672.0
116.90	1277.0
117.80	625.0
118.80	979.0
123.90	151.0
124.80	118.0
127.90	730.0
128.80	380.0
129.90	683.0
130.90	231.0
134.90	309.0
136.80	297.0
140.90	1872.0
141.80	138.0
142.80	1920.0
143.90	182.0
144.80	359.0
145.90	223.0
147.80	372.0
148.80	122.0
149.80	179.0
152.70	115.0
154.90	449.0
156.80	316.0
158.70	213.0
160.80	225.0
170.80	130.0
171.90	1067.0
172.90	1250.0
173.90	132096.0
174.90	9632.0
175.90	125368.0
176.90	8437.0
177.90	271.0

Instrument :
MSVOA_V
ClientSampleId :
BFB280

m/z	Abundance
36.00	1162.0
37.00	6409.0
38.00	5265.0
39.00	2457.0
39.90	596.0
44.00	983.0
45.00	1164.0
47.00	1954.0
47.90	925.0
49.00	5807.0
50.00	27456.0
51.00	8961.0
52.00	406.0
54.90	461.0
56.00	2180.0
57.00	4334.0
58.10	135.0
59.90	1402.0
61.00	7142.0
62.00	7745.0
63.00	5882.0
64.00	706.0
65.00	740.0
66.00	101.0
66.90	466.0
68.00	16424.0
69.00	17096.0
70.00	1329.0
72.00	884.0
73.00	7528.0
74.00	28968.0
75.00	86528.0
76.00	7203.0
76.90	1089.0
77.90	851.0
78.90	5155.0
79.90	1459.0
80.90	5085.0
81.90	1294.0
82.80	137.0
85.80	216.0
86.90	7649.0
87.90	7587.0
90.90	690.0
92.00	5012.0
93.00	7156.0
94.00	20512.0
95.00	171968.0
96.00	11321.0
97.00	285.0
103.90	765.0
104.80	236.0
105.80	818.0
107.00	169.0
109.70	117.0
111.00	155.0
111.70	143.0
114.80	200.0
115.80	720.0
116.90	1220.0
117.80	676.0
118.80	929.0
123.90	161.0
127.90	719.0
128.80	348.0
129.90	653.0
131.00	244.0
134.90	307.0
136.80	288.0
139.80	109.0
140.90	1777.0
141.90	152.0
142.90	1755.0
143.90	141.0
144.80	355.0
145.80	269.0
147.80	399.0
148.80	132.0
149.80	161.0
152.70	107.0
154.80	401.0
156.80	307.0
158.70	251.0
160.80	192.0
170.70	102.0
171.90	1003.0
172.90	1196.0
173.90	127168.0
174.90	9213.0
175.90	121152.0
176.90	8055.0
177.90	236.0

Instrument :
MSVOA_V
ClientSampleId :
BFB280