



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031822\
 Data File : VV025087.D
 Acq On : 18 Mar 2022 09:13
 Operator : SY/MD
 Sample : BFB282
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB282

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Mar 25 23:19:44 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	25109	PASS
75	95	30	80	48.6	77149	PASS
95	95	100	100	100.0	158784	PASS
96	95	5	9	6.5	10317	PASS
173	174	0.00	2	1.0	1133	PASS
174	95	50	120	72.1	114533	PASS
175	174	5	9	7.6	8706	PASS
176	174	95	105	96.0	109984	PASS
177	176	5	10	6.7	7386	PASS

m/z	Abundance
39.90	611.0
43.90	240.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282

m/z	Abundance
36.00	980.0
37.00	5346.0
38.00	5104.0
38.90	2184.0
39.90	536.0
43.90	868.0
45.00	1044.0
46.00	187.0
47.00	1699.0
47.90	769.0
49.00	5540.0
50.00	25816.0
51.00	8044.0
52.00	423.0
54.90	318.0
55.90	2075.0
57.00	3731.0
58.00	140.0
59.90	1180.0
61.00	6499.0
62.00	6832.0
63.00	5942.0
63.90	728.0
65.00	1165.0
66.00	171.0
67.00	401.0
68.00	14610.0
69.00	14856.0
70.00	1275.0
71.90	696.0
73.00	6774.0
74.00	25416.0
75.00	77208.0
76.00	6775.0
76.90	1042.0
78.00	588.0
78.90	4893.0
79.90	1511.0
80.90	5008.0
81.90	1228.0
82.90	103.0
86.00	274.0
86.90	6955.0
87.90	6474.0
90.90	611.0
92.00	4597.0
93.00	7151.0
94.00	18208.0
95.00	157824.0
96.00	10434.0
97.00	315.0
103.80	671.0
104.80	165.0
105.80	662.0
106.80	177.0
110.80	211.0
111.90	115.0
112.80	110.0
114.80	153.0
115.80	530.0
116.90	1147.0
117.80	617.0
118.80	855.0
127.80	536.0
128.80	234.0
129.80	564.0
130.80	190.0
134.80	285.0
136.90	230.0
140.90	1484.0
141.80	212.0
142.90	1663.0
143.80	153.0
144.90	413.0
145.80	227.0
146.80	126.0
147.90	437.0
148.70	117.0
149.80	163.0
152.80	131.0
154.90	374.0
156.90	200.0
158.80	190.0
160.80	209.0
171.00	250.0
171.90	1350.0
172.90	1076.0
173.90	110448.0
175.00	8594.0
175.90	105184.0
176.90	7258.0
177.90	222.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282

m/z	Abundance
36.00	1061.0
37.00	5554.0
38.00	5192.0
39.00	2127.0
39.90	588.0
44.00	909.0
45.00	1039.0
46.00	175.0
47.00	1686.0
47.90	767.0
49.00	5604.0
50.00	26008.0
51.00	8402.0
52.00	359.0
52.90	103.0
55.00	361.0
55.90	2083.0
57.00	3749.0
58.00	163.0
59.90	1246.0
61.00	6682.0
62.00	7067.0
63.00	5987.0
63.90	723.0
65.00	1217.0
66.00	152.0
67.00	375.0
68.00	15155.0
69.00	15748.0
70.00	1269.0
71.90	699.0
73.00	6981.0
74.00	26688.0
75.00	79712.0
76.00	6835.0
76.90	1197.0
78.00	655.0
78.90	5048.0
79.90	1569.0
80.90	5279.0
81.90	1333.0
82.90	122.0
86.10	241.0
86.90	7285.0
88.00	6701.0
90.90	589.0
92.00	4863.0
93.00	7215.0
94.00	18872.0
95.00	164544.0
96.00	10699.0
97.00	306.0
103.80	732.0
104.80	166.0
105.80	661.0
106.80	191.0
110.90	241.0
111.90	157.0
112.90	123.0
114.80	168.0
115.90	564.0
116.80	1108.0
117.80	596.0
118.80	947.0
125.90	104.0
127.80	631.0
128.80	248.0
129.80	671.0
130.80	212.0
134.80	303.0
136.90	222.0
139.90	102.0
140.90	1537.0
141.90	214.0
142.90	1786.0
143.90	144.0
144.80	409.0
145.80	255.0
146.80	110.0
147.90	443.0
148.70	115.0
149.00	109.0
149.80	216.0
152.80	144.0
154.90	386.0
156.90	195.0
158.90	237.0
160.80	209.0
171.00	256.0
171.90	1439.0
173.00	1149.0
173.90	118768.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282

175.00	8969.0
175.90	113896.0
176.90	7658.0
177.90	239.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282

m/z	Abundance
35.90	1068.0
37.00	5128.0
38.00	4785.0
39.00	1981.0
39.90	626.0
44.00	888.0
45.00	1039.0
45.90	137.0
47.00	1601.0
47.90	691.0
49.00	5065.0
50.00	23504.0
51.00	7884.0
52.10	290.0
52.90	122.0
55.00	378.0
55.90	1862.0
57.00	3423.0
58.00	182.0
59.90	1215.0
61.00	6262.0
62.00	6514.0
63.00	5342.0
63.90	647.0
65.00	1089.0
66.00	125.0
67.00	319.0
68.00	14027.0
69.00	14960.0
70.00	1104.0
72.00	693.0
73.00	6428.0
74.00	25064.0
75.00	74528.0
76.00	6206.0
76.90	1135.0
77.90	652.0
78.90	4750.0
79.90	1488.0
80.90	4939.0
81.90	1299.0
82.90	123.0
85.70	140.0
86.00	161.0
86.90	6790.0
88.00	6243.0
90.90	547.0
92.00	4667.0
93.00	6590.0
94.00	17784.0
95.00	153984.0
96.00	9819.0
97.00	214.0
103.80	716.0
104.70	155.0
105.90	608.0
106.90	163.0
110.90	237.0
112.00	165.0
113.00	155.0
114.80	167.0
115.90	543.0
116.80	947.0
117.80	524.0
118.90	944.0
127.80	646.0
128.90	241.0
129.80	664.0
130.80	219.0
134.90	279.0
136.80	210.0
139.90	114.0
140.90	1478.0
141.90	190.0
142.90	1799.0
144.00	129.0
144.80	406.0
145.90	241.0
147.90	369.0
149.80	221.0
152.80	128.0
154.90	357.0
156.90	201.0
158.90	234.0
160.80	164.0
171.00	219.0
171.90	1345.0
172.90	1175.0
173.90	114384.0
175.00	8555.0
175.90	110872.0
176.90	7243.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282

177.90

221.0

Instrument :
MSVOA_V
ClientSampleId :
BFB282