

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031824\
 Data File : VV034688.D
 Acq On : 18 Mar 2024 09:06
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5286

Quant Time: Mar 19 02:18:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	12967	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	11679	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5943	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	7088	0.510	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	4.946	65	4560	0.529	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
7) 1,2-Dichloropropane-d6	5.996	67	4943	0.545	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.254	98	14072	0.523	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3219	0.495	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 98.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	7003	0.484	ug/L	100
6) Trichloroethene	5.853	95	4594	0.502	ug/L	99
9) 1,2-Dibromoethane	8.290	107	2643	0.477	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	2369	0.475	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	486	0.455	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031824\
Data File : VV034688.D
Acq On : 18 Mar 2024 09:06
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5286

Quant Time: Mar 19 02:18:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

