

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031824\
 Data File : VV034691.D
 Acq On : 18 Mar 2024 11:17
 Operator : SY/MD
 Sample : P1715-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGQM2

Quant Time: Mar 19 02:20:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	11090	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.801	117	10293	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.192	152	5360	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	6369	0.536	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.000%			
4) 1,2-Dichloroethane-d4	4.970	65	4127	0.560	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%			
7) 1,2-Dichloropropane-d6	5.996	67	4321	0.540	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.000%			
8) Toluene-d8	7.254	98	11949	0.504	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 100.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3202	0.558	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 112.000%			
Target Compounds						
3) Vinyl chloride	1.281	62	2213	0.179	ug/L	94
6) Trichloroethene	5.829	95	17420	2.161	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031824\
Data File : VV034691.D
Acq On : 18 Mar 2024 11:17
Operator : SY/MD
Sample : P1715-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGQM2

Quant Time: Mar 19 02:20:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

