

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032122\
 Data File : VW025108.D
 Acq On : 21 Mar 2022 10:54
 Operator : SY/MD
 Sample : N2044-04 20X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFY28

Quant Time: Mar 21 12:13:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 19 00:53:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	167795	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	156284	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	72122	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	55832	4.215	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.200%	
7) Chloroethane-d5	1.568	69	43628	4.497	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.000%	
11) 1,1-Dichloroethene-d2	2.105	63	73378	3.427	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.600%	
20) 2-Butanone-d5	3.889	46	93296	58.275	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.540%	
24) Chloroform-d	4.343	84	98938	4.503	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.000%	
26) 1,2-Dichloroethane-d4	5.027	65	43064	4.688	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.800%	
32) Benzene-d6	5.043	84	214945	4.741	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.800%	
36) 1,2-Dichloropropane-d6	6.063	67	63416	4.864	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.200%	
41) Toluene-d8	7.313	98	187392	4.535	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.800%	
43) trans-1,3-Dichloroprop...	7.622	79	19020	4.185	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.800%	
46) 2-Hexanone-d5	8.085	63	77429	49.948	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.900%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	38360	5.002	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	60925	4.990	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.800%	
Target Compounds						Qvalue
16) Methylene chloride	2.507	84	3684	0.284	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025108.D
Acq On : 21 Mar 2022 10:54
Operator : SY/MD
Sample : N2044-04 20X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY28

Quant Time: Mar 21 12:13:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration

