

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032122\
 Data File : VW025110.D
 Acq On : 21 Mar 2022 11:40
 Operator : SY/MD
 Sample : N2044-01 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFY25

Quant Time: Mar 21 12:14:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 19 00:53:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	211895	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	194621	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	94184	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	59456	3.554	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.000%
7) Chloroethane-d5	1.571	69	46149	3.767	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.105	63	78243	2.894	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.800%#
20) 2-Butanone-d5	3.899	46	102615	50.756	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.520%
24) Chloroform-d	4.349	84	107415	3.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
26) 1,2-Dichloroethane-d4	5.034	65	46669	4.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
32) Benzene-d6	5.050	84	230359	4.080	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
36) 1,2-Dichloropropane-d6	6.069	67	65974	4.064	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.200%
41) Toluene-d8	7.317	98	202957	3.944	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	7.625	79	21586	3.814	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	8.092	63	87522	45.338	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.680%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	40946	4.287	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	68251	4.281	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025110.D
Acq On : 21 Mar 2022 11:40
Operator : SY/MD
Sample : N2044-01 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY25

Quant Time: Mar 21 12:14:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration

