

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032122\
 Data File : VW025116.D
 Acq On : 21 Mar 2022 14:01
 Operator : SY/MD
 Sample : N2044-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFY26

Quant Time: Mar 21 14:34:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 19 00:53:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151157	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	142254	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68775	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	51263	4.296	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.000%	
7) Chloroethane-d5	1.568	69	41765	4.779	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.108	63	66473	3.446	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.000%	
20) 2-Butanone-d5	3.912	46	80361	55.720	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.440%	
24) Chloroform-d	4.352	84	93393	4.719	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
26) 1,2-Dichloroethane-d4	5.034	65	40007	4.835	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.600%	
32) Benzene-d6	5.053	84	194748	4.719	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
36) 1,2-Dichloropropane-d6	6.072	67	57558	4.850	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.000%	
41) Toluene-d8	7.317	98	171362	4.556	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.200%	
43) trans-1,3-Dichloroprop...	7.625	79	18390	4.446	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.000%	
46) 2-Hexanone-d5	8.091	63	75142	53.254	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	106.500%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	35008	5.015	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	57296	4.921	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.400%	
Target Compounds						Qvalue
47) Tetrachloroethene	7.992	164	552	0.064	ug/L	# 64
52) Ethylbenzene	9.021	91	2512	0.049	ug/L	97
53) m,p-xylene	9.146	106	2015	0.102	ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025116.D
Acq On : 21 Mar 2022 14:01
Operator : SY/MD
Sample : N2044-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY26

Quant Time: Mar 21 14:34:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration

