

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
 Data File : VV025114.D
 Acq On : 21 Mar 2022 13:14
 Operator : SY/MD
 Sample : N2044-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFY28

Manual IntegrationsAPPROVED

Quant Time: Mar 21 14:33:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 19 00:53:30 2022
 Response via : Initial Calibration

Reviewed By :John Carlone 03/22/2022
 Supervised By :Semsettin Yesilyurt 03/24/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	148958	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	130594	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	58553	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	48898	4.158	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	83.200%		
7) Chloroethane-d5	1.568	69	40368	4.688	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	93.800%		
11) 1,1-Dichloroethene-d2	2.108	63	63255	3.328	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	66.600%		
20) 2-Butanone-d5	3.908	46	76736	53.992	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	107.980%		
24) Chloroform-d	4.352	84	90369	4.633	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	92.600%		
26) 1,2-Dichloroethane-d4	5.037	65	39171	4.804	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	96.000%		
32) Benzene-d6	5.050	84	188690	4.981	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	99.600%		
36) 1,2-Dichloropropane-d6	6.069	67	55571	5.101	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	102.000%		
41) Toluene-d8	7.317	98	160213	4.640	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	92.800%		
43) trans-1,3-Dichloroprop...	7.625	79	16977	4.470	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	89.400%		
46) 2-Hexanone-d5	8.088	63	70436	54.375	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	108.760%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32727	5.107	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	102.200%		
66) 1,2-Dichlorobenzene-d4	11.622	152	47482	4.790	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	95.800%		
Target Compounds						
13) Acetone	2.211	43	7975m	8.390	ug/L	
53) m,p-xylene	9.149	106	1729	0.095	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025114.D
Acq On : 21 Mar 2022 13:14
Operator : SY/MD
Sample : N2044-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY28

Manual IntegrationsAPPROVED

Quant Time: Mar 21 14:33:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration

Reviewed By :John Carlone 03/22/2022
Supervised By :Semsettin Yesilyurt 03/24/2022






Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025114.D
Acq On : 21 Mar 2022 13:14
Operator : SY/MD
Sample : N2044-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY28

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 03/22/2022
Supervised By :Semsettin Yesilyurt 03/24/2022

Quant Time: Mar 21 14:33:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration



TIC: VV025114.D\data.ms

(13) Acetone (T)

2.211min (+ 0.019) 8.39 ug/L m

response 7975

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.10	0.00
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025114.D
Acq On : 21 Mar 2022 13:14
Operator : SY/MD
Sample : N2044-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFY28

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 03/22/2022
Supervised By :Semsettin Yesilyurt 03/24/2022

Quant Time: Apr 07 03:04:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 25 23:19:44 2022
Response via : Initial Calibration



TIC: VV025114.D\data.ms

(13) Acetone (T)

2.211min (+ 0.026) 6.33 ug/L

response 6021

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.10	0.00
0.00	0.00	0.00
0.00	0.00	0.00