

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032122\
 Data File : VW025113.D
 Acq On : 21 Mar 2022 12:51
 Operator : SY/MD
 Sample : N2018-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 21 13:21:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 19 00:53:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151111	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	140892	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62631	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	51652	4.330	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.600%	
7) Chloroethane-d5	1.568	69	42142	4.824	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.400%	
11) 1,1-Dichloroethene-d2	2.108	63	65246	3.384	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.600%	
20) 2-Butanone-d5	3.896	46	84103	58.333	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.660%	
24) Chloroform-d	4.349	84	91740	4.637	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	5.034	65	40176	4.857	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.050	84	195616	4.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.069	67	58097	4.943	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.800%	
41) Toluene-d8	7.317	98	168375	4.520	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
43) trans-1,3-Dichloroprop...	7.625	79	17700	4.320	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.400%	
46) 2-Hexanone-d5	8.088	63	68443	48.975	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.940%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35078	5.073	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55217	5.208	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032122\
Data File : VV025113.D
Acq On : 21 Mar 2022 12:51
Operator : SY/MD
Sample : N2018-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 21 13:21:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 19 00:53:30 2022
Response via : Initial Calibration

