

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032520\
 Data File : VV015144.D
 Acq On : 25 Mar 2020 19:18
 Operator : SY/MD
 Sample : VV0325WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK61

Quant Time: Mar 26 02:14:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032520WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 26 02:00:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	659789	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	626856	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	273690	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	245255	44.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.92%
7) Chloroethane-d5	1.58	69	231424	46.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.62%
11) 1,1-Dichloroethene-d2	2.12	63	376488	34.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.28%
21) 2-Butanone-d5	3.94	46	364262	100.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.62%
24) Chloroform-d	4.38	84	419318	45.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.64%
26) 1,2-Dichloroethane-d4	5.06	65	284362	47.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.78%
32) Benzene-d6	5.08	84	842351	46.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.09	67	265212	46.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.48%
41) Toluene-d8	7.34	98	786853	45.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.10%
43) trans-1,3-Dichloropropene-	7.64	79	142740	45.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.86%
47) 2-Hexanone-d5	8.11	63	284696	99.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.10%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	395088	46.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.56%
64) 1,2-Dichlorobenzene-d4	11.65	152	281075	45.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032520\
Data File : VV015144.D
Acq On : 25 Mar 2020 19:18
Operator : SY/MD
Sample : VV0325WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK61

Quant Time: Mar 26 02:14:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032520WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 26 02:00:02 2020
Response via : Initial Calibration

