

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032520\
 Data File : VV015150.D
 Acq On : 25 Mar 2020 21:44
 Operator : SY/MD
 Sample : L1968-03 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DC9

Quant Time: Mar 26 04:12:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032520WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 26 02:00:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	612767	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	581898	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	263665	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	214279	41.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.64%
7) Chloroethane-d5	1.58	69	204870	44.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.28%
11) 1,1-Dichloroethene-d2	2.12	63	341806	33.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.72%
21) 2-Butanone-d5	3.94	46	314034	93.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.40%
24) Chloroform-d	4.38	84	375299	44.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.32%
26) 1,2-Dichloroethane-d4	5.06	65	256335	46.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.96%
32) Benzene-d6	5.08	84	765517	45.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.26%
36) 1,2-Dichloropropane-d6	6.10	67	241939	45.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.88%
41) Toluene-d8	7.34	98	732155	45.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.32%
43) trans-1,3-Dichloropropene-	7.64	79	115332	39.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.10%
47) 2-Hexanone-d5	8.12	63	242208	90.82	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.82%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	353142	45.04	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.08%
64) 1,2-Dichlorobenzene-d4	11.65	152	266351	45.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.32%

Target Compounds

					Ovalue
33) Benzene	5.13	78	1720673	114.907	ug/L 100
51) Chlorobenzene	8.90	112	2859239	264.853	ug/L 99
62) 1,3-Dichlorobenzene	11.21	146	73835	9.561	ug/L 99
63) 1,4-Dichlorobenzene	11.30	146	1300087	165.249	ug/L 98
65) 1,2-Dichlorobenzene	11.67	146	880667	113.031	ug/L 99
68) 1,2,4-trichlorobenzene	13.28	180	109913	22.432	ug/L 99
70) 1,2,3-Trichlorobenzene	13.77	180	9396	1.930	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032520\
Data File : VV015150.D
Acq On : 25 Mar 2020 21:44
Operator : SY/MD
Sample : L1968-03 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DC9

Quant Time: Mar 26 04:12:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032520WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 26 02:00:02 2020
Response via : Initial Calibration

