

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025199.D
 Acq On : 26 Mar 2022 15:38
 Operator : SY/MD
 Sample : N2100-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 28 02:01:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	149496	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	143993	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62548	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	71651	5.131	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.600%		
7) Chloroethane-d5	1.568	69	61388	5.238	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 104.800%		
11) 1,1-Dichloroethene-d2	2.108	63	115138	3.806	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 76.200%		
20) 2-Butanone-d5	3.918	46	61548	44.247	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 88.500%		
24) Chloroform-d	4.352	84	89538	4.913	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.200%		
26) 1,2-Dichloroethane-d4	5.037	65	41418	5.142	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.800%		
32) Benzene-d6	5.050	84	185435	5.090	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.800%		
36) 1,2-Dichloropropane-d6	6.069	67	53669	5.257	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 105.200%		
41) Toluene-d8	7.317	98	164085	4.911	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 98.200%		
43) trans-1,3-Dichloroprop...	7.625	79	15526	4.662	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 93.200%		
46) 2-Hexanone-d5	8.092	63	59107	49.138	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 98.280%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33347	4.899	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 98.000%		
66) 1,2-Dichlorobenzene-d4	11.622	152	53783	5.354	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 107.000%		
Target Compounds						
16) Methylene chloride	2.510	84	4367	0.370	ug/L	96
25) Chloroform	4.375	83	2641	0.132	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025199.D
Acq On : 26 Mar 2022 15:38
Operator : SY/MD
Sample : N2100-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 28 02:01:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 01:44:37 2022
Response via : Initial Calibration

