

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025201.D
 Acq On : 26 Mar 2022 16:25
 Operator : SY/MD
 Sample : N2083-06DL 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW5Z9DL

Quant Time: Mar 28 02:01:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151141	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	144320	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63027	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	75714	5.363	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 107.200%		
7) Chloroethane-d5	1.568	69	65587	5.535	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 110.800%		
11) 1,1-Dichloroethene-d2	2.108	63	115499	3.777	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 75.600%		
20) 2-Butanone-d5	3.918	46	58612	41.678	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 83.360%		
24) Chloroform-d	4.349	84	94528	5.131	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 102.600%		
26) 1,2-Dichloroethane-d4	5.034	65	45421	5.577	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 111.600%		
32) Benzene-d6	5.050	84	194049	5.315	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.200%		
36) 1,2-Dichloropropane-d6	6.072	67	56150	5.487	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 109.800%		
41) Toluene-d8	7.317	98	167181	4.992	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 99.800%		
43) trans-1,3-Dichloroprop...	7.625	79	16973	5.085	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 101.600%		
46) 2-Hexanone-d5	8.092	63	62811	52.099	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 104.200%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35757	5.241	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 104.800%		
66) 1,2-Dichlorobenzene-d4	11.622	152	56149	5.547	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 111.000%		
Target Compounds						
						Qvalue
16) Methylene chloride	2.507	84	13923	1.165	ug/L	94
19) 1,1-Dichloroethane	3.198	63	4047	0.222	ug/L #	93
25) Chloroform	4.375	83	7316	0.362	ug/L	92
29) 1,1,1-Trichloroethane	4.606	97	27241	1.561	ug/L	98
34) Trichloroethene	5.918	95	14464	1.244	ug/L	91
47) Tetrachloroethene	7.976	164	26887	3.137	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025201.D
Acq On : 26 Mar 2022 16:25
Operator : SY/MD
Sample : N2083-06DL 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW5Z9DL

Quant Time: Mar 28 02:01:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 01:44:37 2022
Response via : Initial Calibration

