

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW032622\  
 Data File : VW025205.D  
 Acq On : 26 Mar 2022 17:59  
 Operator : SY/MD  
 Sample : N2082-16  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 EW615

Quant Time: Mar 28 02:02:42 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Mar 28 01:44:37 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114            | 154363   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117            | 150597   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 63187    | 5.000  | ug/L     | 0.00     |
|                               |        |                |          |        |          |          |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.307  | 65             | 66179    | 4.590  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 91.800%  |          |
| 7) Chloroethane-d5            | 1.568  | 69             | 59762    | 4.938  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 98.800%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63             | 104388   | 3.342  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 66.800%  |          |
| 20) 2-Butanone-d5             | 3.899  | 46             | 74707    | 52.014 | ug/L     | -0.02    |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 104.020% |          |
| 24) Chloroform-d              | 4.352  | 84             | 95563    | 5.079  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 101.600% |          |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65             | 43722    | 5.257  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 105.200% |          |
| 32) Benzene-d6                | 5.050  | 84             | 195239   | 5.124  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 102.400% |          |
| 36) 1,2-Dichloropropane-d6    | 6.069  | 67             | 54270    | 5.083  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 101.600% |          |
| 41) Toluene-d8                | 7.317  | 98             | 170322   | 4.874  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 97.400%  |          |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79             | 16061    | 4.611  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 92.200%  |          |
| 46) 2-Hexanone-d5             | 8.088  | 63             | 58490    | 46.493 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 92.980%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214 | 84             | 33932    | 4.766  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 95.400%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152            | 55332    | 5.452  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 109.000% |          |
|                               |        |                |          |        |          |          |
| Target Compounds              |        |                |          |        |          | Qvalue   |
| 25) Chloroform                | 4.375  | 83             | 6186     | 0.299  | ug/L     | 98       |
| 38) Bromodichloromethane      | 6.526  | 83             | 1516     | 0.117  | ug/L #   | 83       |
| 42) Toluene                   | 7.394  | 91             | 10277    | 0.218  | ug/L     | 99       |
| -----                         |        |                |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV032622\  
Data File : VV025205.D  
Acq On : 26 Mar 2022 17:59  
Operator : SY/MD  
Sample : N2082-16  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
EW615

Quant Time: Mar 28 02:02:42 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Mon Mar 28 01:44:37 2022  
Response via : Initial Calibration

