

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025210.D
 Acq On : 26 Mar 2022 19:56
 Operator : SY/MD
 Sample : N2097-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD3

Quant Time: Mar 28 02:03:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	153771	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	149660	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64402	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	63821	4.443	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.800%	
7) Chloroethane-d5	1.568	69	56809	4.712	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.200%	
11) 1,1-Dichloroethene-d2	2.108	63	102340	3.289	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.800%	
20) 2-Butanone-d5	3.905	46	66704	46.621	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.240%	
24) Chloroform-d	4.349	84	92573	4.939	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.800%	
26) 1,2-Dichloroethane-d4	5.034	65	41624	5.024	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.400%	
32) Benzene-d6	5.053	84	185658	4.904	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
36) 1,2-Dichloropropane-d6	6.069	67	55469	5.227	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.600%	
41) Toluene-d8	7.317	98	161668	4.655	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
43) trans-1,3-Dichloroprop...	7.625	79	14701	4.247	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.000%	
46) 2-Hexanone-d5	8.088	63	58370	46.688	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.380%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32741	4.628	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	53887	5.210	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	
Target Compounds						
9) Trichlorofluoromethane	1.751	101	2370	0.087	ug/L	95
10) 1,1,2-Trichloro-1,2,2-...	2.117	101	32174	1.980	ug/L	99
17) Methyl tert-butyl Ether	2.773	73	11148	0.569	ug/L #	95
19) 1,1-Dichloroethane	3.195	63	2686	0.145	ug/L #	94
25) Chloroform	4.378	83	49062	2.384	ug/L	99
29) 1,1,1-Trichloroethane	4.609	97	26231	1.450	ug/L	100
34) Trichloroethene	5.915	95	30342	2.517	ug/L	97
47) Tetrachloroethene	7.976	164	2203	0.248	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
 Data File : VV025210.D
 Acq On : 26 Mar 2022 19:56
 Operator : SY/MD
 Sample : N2097-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD3

Quant Time: Mar 28 02:03:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

