

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025215.D
 Acq On : 26 Mar 2022 21:53
 Operator : SY/MD
 Sample : N2097-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD8

Quant Time: Mar 28 02:04:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	149501	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142329	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60096	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	71883	5.147	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.000%
7) Chloroethane-d5	1.568	69	61082	5.212	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.200%
11) 1,1-Dichloroethene-d2	2.108	63	111490	3.686	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.800%
20) 2-Butanone-d5	3.925	46	67390	48.446	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.900%
24) Chloroform-d	4.349	84	94750	5.199	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.034	65	42553	5.283	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.050	84	189844	5.272	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
36) 1,2-Dichloropropane-d6	6.069	67	54798	5.430	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.600%
41) Toluene-d8	7.317	98	164846	4.991	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
43) trans-1,3-Dichloroprop...	7.625	79	15602	4.739	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.092	63	59925	50.400	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.800%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	34701	5.157	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	55342	5.734	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.600%
Target Compounds						
17) Methyl tert-butyl Ether	2.773	73	2210	0.116	ug/L	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025215.D
Acq On : 26 Mar 2022 21:53
Operator : SY/MD
Sample : N2097-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD8

Quant Time: Mar 28 02:04:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 01:44:37 2022
Response via : Initial Calibration

