

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
 Data File : VV025223.D
 Acq On : 27 Mar 2022 01:24
 Operator : SY/MD
 Sample : VV0326WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK326

Quant Time: Mar 28 03:13:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151321	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	145911	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62026	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	70137	4.962	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	99.200%	
7) Chloroethane-d5	1.571	69	62011	5.227	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.600%	
11) 1,1-Dichloroethene-d2	2.111	63	111356	3.637	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.800%	
20) 2-Butanone-d5	3.915	46	62636	44.487	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	88.980%	
24) Chloroform-d	4.352	84	95078	5.154	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.000%	
26) 1,2-Dichloroethane-d4	5.034	65	44349	5.439	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.800%	
32) Benzene-d6	5.053	84	192276	5.209	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.200%	
36) 1,2-Dichloropropane-d6	6.072	67	54823	5.299	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.000%	
41) Toluene-d8	7.317	98	166674	4.923	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.400%	
43) trans-1,3-Dichloroprop...	7.625	79	15566	4.612	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.200%	
46) 2-Hexanone-d5	8.091	63	59111	48.495	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.000%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34714	5.032	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55048	5.526	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.600%	
Target Compounds					Qvalue	
25) Chloroform	4.387	83	3036	0.150	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025223.D
Acq On : 27 Mar 2022 01:24
Operator : SY/MD
Sample : VV0326WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK326

Quant Time: Mar 28 03:13:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

