

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025228.D
 Acq On : 27 Mar 2022 03:21
 Operator : SY/MD
 Sample : N2097-20
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAFO

Quant Time: Mar 28 03:47:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	147776	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	146139	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61188	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62741	4.545	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.000%	
7) Chloroethane-d5	1.568	69	57182	4.936	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.800%	
11) 1,1-Dichloroethene-d2	2.111	63	102432	3.426	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.600%	
20) 2-Butanone-d5	3.915	46	65482	47.624	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.240%	
24) Chloroform-d	4.352	84	91279	5.067	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.400%	
26) 1,2-Dichloroethane-d4	5.034	65	42904	5.388	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.800%	
32) Benzene-d6	5.050	84	185030	5.005	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.000%	
36) 1,2-Dichloropropane-d6	6.069	67	53459	5.159	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.200%	
41) Toluene-d8	7.317	98	158759	4.682	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.600%	
43) trans-1,3-Dichloroprop...	7.625	79	14855	4.395	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.800%	
46) 2-Hexanone-d5	8.092	63	57045	46.727	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.460%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32667	4.728	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	53897	5.484	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.600%	
Target Compounds						
9) Trichlorofluoromethane	1.757	101	2469	0.094	ug/L	94
17) Methyl tert-butyl Ether	2.773	73	40477	2.150	ug/L	98
19) 1,1-Dichloroethane	3.195	63	5560	0.312	ug/L	99
25) Chloroform	4.378	83	10711	0.542	ug/L	100
29) 1,1,1-Trichloroethane	4.613	97	1015	0.057	ug/L	89
34) Trichloroethene	5.924	95	3829	0.325	ug/L	85
47) Tetrachloroethene	7.976	164	12973	1.495	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025228.D
Acq On : 27 Mar 2022 03:21
Operator : SY/MD
Sample : N2097-20
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AFO

Quant Time: Mar 28 03:47:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

