

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025229.D
 Acq On : 27 Mar 2022 03:44
 Operator : SY/MD
 Sample : N2110-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0PB1

Quant Time: Mar 28 03:48:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	147088	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142797	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61147	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	60675	4.416	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.571	69	54929	4.763	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.111	63	97304	3.270	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.400%
20) 2-Butanone-d5	3.921	46	56594	41.352	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.700%
24) Chloroform-d	4.352	84	84880	4.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.037	65	38777	4.893	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.053	84	170711	4.725	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	6.069	67	49435	4.883	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.317	98	149769	4.520	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
43) trans-1,3-Dichloroprop...	7.625	79	13292	4.024	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.400%
46) 2-Hexanone-d5	8.092	63	56524	47.384	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.760%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30340	4.494	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	50816	5.174	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.114	96	2117	0.143	ug/L #	1
17) Methyl tert-butyl Ether	2.777	73	3250	0.173	ug/L	96
25) Chloroform	4.378	83	9282	0.471	ug/L	86
29) 1,1,1-Trichloroethane	4.609	97	829	0.048	ug/L #	87
47) Tetrachloroethene	7.973	164	441996	52.116	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025229.D
Acq On : 27 Mar 2022 03:44
Operator : SY/MD
Sample : N2110-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0PB1

Quant Time: Mar 28 03:48:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

