

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025231.D
 Acq On : 27 Mar 2022 04:31
 Operator : SY/MD
 Sample : N2110-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0PB3

Quant Time: Mar 28 03:48:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	145592	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	138341	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59453	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62157	4.570	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.400%
7) Chloroethane-d5	1.571	69	56237	4.927	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.600%
11) 1,1-Dichloroethene-d2	2.111	63	101023	3.429	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.600%
20) 2-Butanone-d5	3.908	46	64096	47.315	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.620%
24) Chloroform-d	4.352	84	91563	5.159	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
26) 1,2-Dichloroethane-d4	5.034	65	42212	5.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.053	84	182899	5.226	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
36) 1,2-Dichloropropane-d6	6.072	67	52660	5.369	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.400%
41) Toluene-d8	7.317	98	159229	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.200%
43) trans-1,3-Dichloroprop...	7.625	79	14635	4.574	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.092	63	65376	56.570	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.140%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32068	4.903	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	54291	5.686	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.800%
Target Compounds						
					Qvalue	
17) Methyl tert-butyl Ether	2.780	73	4612	0.249	ug/L	# 90
22) cis-1,2-Dichloroethene	3.925	96	1239	0.116	ug/L	86
25) Chloroform	4.378	83	11204	0.575	ug/L	95
34) Trichloroethene	5.921	95	5235	0.470	ug/L	94
47) Tetrachloroethene	7.979	164	5354064	651.637	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025231.D
Acq On : 27 Mar 2022 04:31
Operator : SY/MD
Sample : N2110-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0PB3

Quant Time: Mar 28 03:48:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

