

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025232.D
 Acq On : 27 Mar 2022 04:55
 Operator : SY/MD
 Sample : N2110-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0PB4

Quant Time: Mar 28 03:48:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	145295	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	141977	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	59015	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	65795	4.848	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.000%
7) Chloroethane-d5	1.571	69	57300	5.030	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%
11) 1,1-Dichloroethene-d2	2.108	63	104268	3.547	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	3.912	46	63642	47.076	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.160%
24) Chloroform-d	4.352	84	93553	5.282	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
26) 1,2-Dichloroethane-d4	5.034	65	41569	5.310	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.200%
32) Benzene-d6	5.050	84	186668	5.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
36) 1,2-Dichloropropane-d6	6.072	67	53379	5.303	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.317	98	161018	4.887	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
43) trans-1,3-Dichloroprop...	7.625	79	14741	4.489	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.800%
46) 2-Hexanone-d5	8.091	63	59320	50.015	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.040%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33686	5.019	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	55078	5.811	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.200%
Target Compounds						
					Qvalue	
17) Methyl tert-butyl Ether	2.777	73	3364	0.182	ug/L #	90
25) Chloroform	4.381	83	6043	0.311	ug/L	91
29) 1,1,1-Trichloroethane	4.609	97	1042	0.061	ug/L #	48
47) Tetrachloroethene	7.972	164	113578	13.469	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025232.D
Acq On : 27 Mar 2022 04:55
Operator : SY/MD
Sample : N2110-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0PB4

Quant Time: Mar 28 03:48:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

